

Certificate of Calibration

Customer :

Description of Equipment : Digital Multimeter
Model Number : 3458A
Manufacturer : Hewlett Packard
Serial Number :
Control Number :
Receipt Date :
Calibration Date :
Work Order Number :
Certificate No. :
Page Number : 1 of 12

We certify that the above mentioned measuring instrument was calibrated in accordance with our quality control system as the ISO/IEC 17025 : 2005 and all the standard equipments used for calibration were traceable to the International of System Unit (SI Unit)

Issued Date :
Calibrated by :
Checked by :

Approved by : Mr. Yanyong Pithong / Laboratory Manager
 Mr. Satoshi Fujiwara / Head of Laboratory

Approved Signatory

Calibration Report

Certificate No.:

Page : 2 of 12

Environment Conditions

The calibration was carried out in an ambient of temperature $(23 \pm 3)^{\circ}\text{C}$ and relative humidity $(55 \pm 15)\%\text{RH}$ and accomplished in an ambient environment controlled in the laboratory.

Standard Equipment Used

Equip. No.	Equipment Name	Maker & Model	Serial No.	Cert. No.	Due Date
SE-00009	Multi-Function Calibrator	Fluke & 5720A	6960202	EL-0056/08	25-Mar-09
SE-99010	Amplifier	Fluke & 5725A	6485001	EL-0057/08	25-Mar-09
SE-04179	AC Measurement System	Fluke & 5790A+WB	5510033	EL-0003/08	14-Feb-10
SE-01020	Digital Multimeter	Agilent & 3458A-002	2823A27401	EL-0193/08	10-Sep-09
SE-99049	Standard Resistor : 1Ω	Fluke & 742A-1	6330024	EL-0087/08	31-Mar-09
SE-07252	Standard Resistance	IET Labs, Inc. & SRX-10	H1-0646479	108-3030	23-Feb-09
SE-07253	Standard Resistance	IET Labs, Inc. & SRX-100	H1-0646480	108-3031	23-Feb-09
SE-07254	Standard Resistance	IET Labs, Inc. & SRX-1Y	H1-0646481	108-3032	26-Feb-09
SE-07257	Standard Resistance	IET Labs, Inc. & SKX-100M	H1-0710500	108-3033	28-Feb-09
SE-07251	Standard Resistance	IET Labs, Inc. & SKX-1M	H1-0705312	108-3034	04-Mar-09
SE-07255	Standard Resistance	IET Labs, Inc. & SRX-10M	H1-0710498	108-3035	03-Mar-09
SE-99025	DC/AC Shunt	Guiloline & 7320	63834	EL-0216/06	09-Feb-09
SE-99095	Synthesized Func/Sweep Gen.	HP & 3325B	2847A09782	E3080092	02-May-09

Traceability

This certificate is traceable to the International System of Unit (SI Unit) maintained at;

- National Institute of Metrology (NIMT), Thailand for dc voltage standards (0 to 1100V).
- National Institute of Metrology (NIMT), Thailand for ac voltage standards (0.22mV to 1100V, 10Hz to 10MHz).
- National Institute of Metrology (NIMT), Thailand for dc current standards (0 to 200A).
- National Institute of Metrology (NIMT), Thailand for ac current standards (9 μA to 20.5A, 10Hz to 100kHz).
- National Institute of Metrology (NIMT), Thailand for resistance standards (0 to 1T Ω).
- National Institute of Metrology (NIMT), Thailand for time and frequency standards.

Calibration Report

Certificate No.:

Page : 3 of 12

Uncertainty of Measurement

The uncertainty of measurement evaluation has been carried out by using the methodology in the "Expression of Uncertainty and Confidence in Measurement" (M3003). These uncertainties are based on a standard uncertainty multiplied by a coverage factor $k = 2$

Calibration Method

The Agilent/HP 3458A Multimeter (Unit Under Calibration : UUC) has been calibrated in accordance with the in-house method of the laboratory and the calibration procedures described in the Agilent Technologies 3458A Multimeter Calibration Manual P/N 03458-

The identification of the laboratory's calibration procedures employed are E-DCV-M-0004-GT, E-ACV-M-0004-GT, E-DCA-M-0004-GT, E-ACA-M-0001-GT, E-ACA-M-0004-GT and E-DCK M-0001-GT.

Calibration Results

The following results were the measurement results without any adjustment and were applied on the calibrated item and found accurate as shown on the date and place of calibration only.

UUC Temperature Check

The 3458A's displayed temperature last adjust (CAL 0) was	38.00 °C
The 3458A's displayed temperature last adjust (CAL 10V) was	38.00 °C
The 3458A's displayed temperature last adjust (CAL 10kΩ) was	38.00 °C

Calibration Report

Certificate No.:

Page : 4 of 12

Calibration Results (Cont.)

DC Voltage Performance Tests

Test Conditions:

The 3458A was turned on at least 48-hours before calibration.

The 3458A's displayed internal temperature (TEMP?) was 38.0 °C

The 3458A was in the default/power-on state with the following exceptions:

Function: DCV

Ranges: Manual selection

NPLC 100 (Integration time 100 power line cycles)

NDIG 8 (Resolution 8 digits)

The 3458A's front input terminals were used.

The Guard switch was in "Open" position.

The 3458A's "Guard" and "Lo" terminals were at Earth potential.

The 3458A was allowed to settle for 30 minute after "AUTOCAL" was performed.

DC Voltage Function Offset Test (MATH OFF, Shorting time > 5 min.)

UUC Range	Applied Value	UUC Reading	Uncertainty (±)	UUC Error	1Year Spec. (±)
100 mV	(Short) 0 mV	0.000000 mV	0.00040 mV	0.00000 mV	0.00106 mV
1 V	(Short) 0 V	0.0000000 V	0.00040 mV	0.00000 mV	0.00106 mV
10 V	(Short) 0 V	0.0000000 V	0.00040 mV	0.00000 mV	0.0023 mV
100 V	(Short) 0 V	0.000000 V	0.0010 mV	0.000 mV	0.036 mV
1000 V	(Short) 0 V	0.00000 V	0.010 mV	0.00 mV	0.10 mV

DC Voltage Function Gain Test

(Performed "MATH NULL" with the input terminals shorted, prior to starting calibration on each range)

UUC Range	Applied Value	UUC Reading	Uncertainty (±)	UUC Error	1Year Spec. (±)
100 mV	100.000097 mV	100.00010 mV	11 •10 ⁻⁶	0.00000 mV	0.00188 mV
100 mV	-99.999708 mV	-99.99971 mV	10 •10 ⁻⁶	0.00000 mV	0.00188 mV
1 V	0.9999979 V	0.9999979 V	5.0 •10 ⁻⁶	0.00000 mV	0.00740 mV
1 V	-0.9999958 V	-0.9999958 V	5.0 •10 ⁻⁶	0.00000 mV	0.00740 mV
10 V	9.999996 V	9.999996 V	5.0 •10 ⁻⁶	0.0000 mV	0.0624 mV
10 V	-10.000089 V	-10.000089 V	5.1 •10 ⁻⁶	0.0000 mV	0.0624 mV
100 V	99.999888 V	99.999888 V	5.0 •10 ⁻⁶	0.000 mV	0.853 mV
100 V	-99.99989 V	-99.99989 V	5.0 •10 ⁻⁶	0.000 mV	0.853 mV
1000 V	1.0000053 kV	1.0000053 kV	6.1 •10 ⁻⁶	0.00 mV	19.34 mV
1000 V	-1.0000116 kV	-1.0000116 kV	6.1 •10 ⁻⁶	0.00 mV	19.34 mV

Calibration Report

Certificate No.:

Page : 5 of 12

Calibration Results (Cont.)

Analog AC Voltage Performance Tests (Mode ANA)

Test Conditions:

The 3458A was turned on at least 48-hours before calibration.

The 3458A's displayed internal temperature (TEMP?) was 38.0 °C

The 3458A was in the default/power-on state with the following exceptions:

Function: ACV

Ranges: Manual selection

SETACV ANA (Analog Mode)

ACBAND 10, 2E6 (10Hz~2MHz)

LFILTER ON (Low-pass filter ON)

NPLC 100 (Integration time 100 power line cycles)

The 3458A's front input terminals were used.

The Guard switch was in "Open" position.

The 3458A's "Guard" and "Lo" terminals were at Earth potential.

The 3458A was performed "AUTOCAL" prior to starting the measurements.

UUC Range	Applied Value	Freq.	UUC Reading	Uncertainty (±)	UUC Error	1 Year Spec. (±)
10 mV	9.998310 mV	1 kHz	9.99231 mV	400 • 10 ⁻⁶	0.000000 mV	0.027020 mV
100 mV	100.00218 mV	1 kHz	100.0022 mV	160 • 10 ⁻⁶	0.000020 mV	0.030200 mV
1 V	1.0000212 V	1 kHz	1.000021 V	160 • 10 ⁻⁶	0.000000 V	0.000302 V
10 V	1.0000212 V	1 kHz	1.00002 V	160 • 10 ⁻⁶	-0.000001 V	0.001202 V
10 V	1.0000298 V	50 kHz	1.00003 V	160 • 10 ⁻⁶	0.000000 V	0.005502 V
10 V	1.0011024 V	1 MHz	1.00110 V	1598 • 10 ⁻⁶	-0.000002 V	0.250057 V
10 V	10.000469 V	10 Hz	10.00047 V	81 • 10 ⁻⁶	0.000001 V	0.042022 V
10 V	10.000192 V	20 Hz	10.00019 V	81 • 10 ⁻⁶	-0.000002 V	0.042021 V
10 V	10.000145 V	40 Hz	10.00015 V	81 • 10 ⁻⁶	0.000005 V	0.017020 V
10 V	10.000121 V	100 Hz	10.00012 V	81 • 10 ⁻⁶	-0.000001 V	0.007020 V
10 V	10.000114 V	1 kHz	10.00011 V	81 • 10 ⁻⁶	-0.000004 V	0.003020 V
10 V	10.000121 V	20 kHz	10.00012 V	81 • 10 ⁻⁶	-0.000001 V	0.003020 V
10 V	10.000123 V	50 kHz	10.00012 V	81 • 10 ⁻⁶	-0.000003 V	0.019020 V
10 V	10.000109 V	100 kHz	10.00011 V	81 • 10 ⁻⁶	0.000001 V	0.068021 V
10 V	10.000696 V	300 kHz	10.00070 V	610 • 10 ⁻⁶	0.000004 V	0.360041 V
10 V	10.001960 V	500 kHz	10.00196 V	710 • 10 ⁻⁶	0.000000 V	0.360079 V
10 V	10.006970 V	1 MHz	10.00697 V	2598 • 10 ⁻⁶	0.000000 V	0.700369 V
100 V	99.99995 V	1 kHz	100.0000 V	80 • 10 ⁻⁶	0.000050 V	0.040200 V
1000 V	600.0046 V	1 kHz	600.005 V	82 • 10 ⁻⁶	0.000400 V	0.561203 V

Calibration Report

Certificate No.:

Page : 6 of 12

Calibration Results (Cont.)

Synchronously Sub-sampled AC Voltage Performance Tests (Mode SYNC)

Test Conditions:

The 3458A was turned on at least 48-hours before calibration.

The 3458A's displayed internal temperature (TEMP?) was 38.0 °C

The 3458A was in the default/power-on state with the following exceptions:

Function: ACV

Ranges: Manual selection

SETACV SYNC (Synchronously Sub-sampled Mode)

ACBAND 10, 2E6 (10Hz~2MHz)

LFILTER ON (Low-pass filter ON)

NPLC 100 (Integration time 100 power line cycles)

RES .002 (AC reading resolution 6 digits)

The 3458A's front input terminals were used.

The Guard switch was in "Open" position.

The 3458A's "Guard" and "Lo" terminals were at Earth potential.

The 3458A was performed "AUTOCAL" prior to starting the measurements.

UUC Range	Applied Value	Freq.	UUC Reading	Uncertainty (\pm)	UUC Error	1Year Spec. (\pm)
10 mV	9.998810 mV	10 Hz	9.99881 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.00602 mV
10 mV	9.998610 mV	20 Hz	9.99861 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.00602 mV
10 mV	9.998490 mV	40 Hz	9.99849 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.00602 mV
10 mV	9.998400 mV	100 Hz	9.99840 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.00312 mV
10 mV	9.998310 mV	1 kHz	9.99831 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.00312 mV
10 mV	9.999530 mV	10 kHz	9.99953 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.00412 mV
10 mV	9.999760 mV	20 kHz	9.99976 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.00412 mV
10 mV	9.999860 mV	50 kHz	9.99986 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.01112 mV
10 mV	9.996499 mV	100 kHz	9.99650 mV	$400 \cdot 10^{-6}$	0.00000 mV	0.05110 mV
10 mV	9.995949 mV	300 kHz	9.99595 mV	$800 \cdot 10^{-6}$	0.00000 mV	0.40186 mV
10 mV	9.994212 mV	500 kHz	9.99421 mV	$800 \cdot 10^{-6}$	0.00000 mV	n/a
10 mV	9.982696 mV	** 1 MHz	9.98270 mV	$2003 \cdot 10^{-6}$	0.00000 mV	0.12481 mV

Note: (**) ACBAND 1E6, 10E6 (1MHz~10MHz), LFILTER OFF (Low-pass filter OFF)

Calibration Report

Certificate No.:

Page : 7 of 12

Calibration Results (Cont.)

Synchronously Sub-sampled AC Voltage Performance Tests (Mode SYNC)

Cont.

UUC Range	Applied Value	Freq.	UUC Reading	Uncertainty (\pm)	UUC Error	1Year Spec. (\pm)
100 mV	100.00594 mV	10 Hz	100.0059 mV	$160 \cdot 10^{-6}$	0.0000 mV	0.0112 mV
100 mV	100.00302 mV	20 Hz	100.0030 mV	$160 \cdot 10^{-6}$	0.0000 mV	0.0112 mV
100 mV	100.00230 mV	40 Hz	100.0023 mV	$160 \cdot 10^{-6}$	0.0000 mV	0.0112 mV
100 mV	100.00221 mV	100 Hz	100.0022 mV	$160 \cdot 10^{-6}$	0.0000 mV	0.0092 mV
100 mV	100.00218 mV	1 kHz	100.0022 mV	$160 \cdot 10^{-6}$	0.0000 mV	0.0092 mV
100 mV	100.00264 mV	10 kHz	100.0026 mV	$160 \cdot 10^{-6}$	0.0000 mV	0.0162 mV
100 mV	100.00196 mV	20 kHz	100.0020 mV	$160 \cdot 10^{-6}$	0.0000 mV	0.0162 mV
100 mV	99.99988 mV	50 kHz	99.9999 mV	$160 \cdot 10^{-6}$	0.0000 mV	0.0322 mV
100 mV	99.99200 mV	100 kHz	99.9920 mV	$150 \cdot 10^{-6}$	0.0000 mV	0.0822 mV
100 mV	99.96077 mV	300 kHz	99.9608 mV	$600 \cdot 10^{-6}$	0.0000 mV	0.3101 mV
100 mV	99.86944 mV	500 kHz	99.8694 mV	$701 \cdot 10^{-6}$	0.0000 mV	1.0089 mV
100 mV	99.49188 mV	1 MHz	99.4919 mV	$2513 \cdot 10^{-6}$	0.0000 mV	1.0051 mV
1 V	1.0000577 V	10 Hz	1.000058 V	$160 \cdot 10^{-6}$	0.000000 V	0.000112 V
1 V	1.0000281 V	20 Hz	1.000028 V	$160 \cdot 10^{-6}$	0.000000 V	0.000112 V
1 V	1.0000225 V	40 Hz	1.000023 V	$160 \cdot 10^{-6}$	0.000001 V	0.000112 V
1 V	1.0000214 V	100 Hz	1.000021 V	$160 \cdot 10^{-6}$	0.000000 V	0.000092 V
1 V	1.0000212 V	1 kHz	1.000021 V	$160 \cdot 10^{-6}$	0.000000 V	0.000092 V
1 V	1.0000214 V	10 kHz	1.000021 V	$160 \cdot 10^{-6}$	0.000000 V	0.000162 V
1 V	1.0000250 V	20 kHz	1.000025 V	$160 \cdot 10^{-6}$	0.000000 V	0.000162 V
1 V	1.0000298 V	50 kHz	1.000030 V	$160 \cdot 10^{-6}$	0.000000 V	0.000322 V
1 V	1.0000347 V	100 kHz	1.000035 V	$260 \cdot 10^{-6}$	0.000000 V	0.000822 V
1 V	1.0018950 V	300 kHz	1.001895 V	$759 \cdot 10^{-6}$	0.000000 V	0.003108 V
1 V	1.0004184 V	500 kHz	1.000418 V	$760 \cdot 10^{-6}$	0.000000 V	0.010106 V
1 V	1.0011024 V	1 MHz	1.001102 V	$1598 \cdot 10^{-6}$	0.000000 V	0.010113 V
10 V	10.000469 V	10 Hz	10.00047 V	$81 \cdot 10^{-6}$	0.00000 V	0.00112 V
10 V	10.000192 V	20 Hz	10.00019 V	$81 \cdot 10^{-6}$	0.00000 V	0.00112 V
10 V	10.000145 V	40 Hz	10.00015 V	$81 \cdot 10^{-6}$	0.00000 V	0.00112 V
10 V	10.000121 V	100 Hz	10.00012 V	$81 \cdot 10^{-6}$	0.00000 V	0.00092 V
10 V	10.000114 V	1 kHz	10.00011 V	$81 \cdot 10^{-6}$	0.00000 V	0.00092 V
10 V	10.000118 V	10 kHz	10.00012 V	$81 \cdot 10^{-6}$	0.00000 V	0.00162 V
10 V	10.000121 V	20 kHz	10.00012 V	$81 \cdot 10^{-6}$	0.00000 V	0.00162 V
10 V	10.000123 V	50 kHz	10.00012 V	$81 \cdot 10^{-6}$	0.00000 V	0.00322 V
10 V	10.000109 V	100 kHz	10.00011 V	$81 \cdot 10^{-6}$	0.00000 V	0.00822 V
10 V	10.000696 V	300 kHz	10.00070 V	$610 \cdot 10^{-6}$	0.00000 V	0.03102 V
10 V	10.001960 V	500 kHz	10.00196 V	$710 \cdot 10^{-6}$	0.00000 V	0.10104 V
10 V	10.006970 V	1 MHz	10.00697 V	$2598 \cdot 10^{-6}$	0.00000 V	0.10109 V

Calibration Report

Certificate No.:

Page : 8 of 12

Calibration Results (Cont.)

Synchronously Sub-sampled AC Voltage Performance Tests (Mode SYNC)

Cont.

UUC Range	Applied Value	Freq.	UUC Reading	Uncertainty (\pm)	UUC Error	1 Year Spec. (\pm)
100 V	100.00320 V	10 Hz	100.0032 V	$81 \cdot 10^{-6}$	0.0000 V	0.0242 V
100 V	100.00011 V	20 Hz	100.0001 V	$81 \cdot 10^{-6}$	0.0000 V	0.0242 V
100 V	99.99965 V	40 Hz	99.9997 V	$80 \cdot 10^{-6}$	0.0001 V	0.0242 V
100 V	99.99958 V	100 Hz	99.9996 V	$80 \cdot 10^{-6}$	0.0000 V	0.0222 V
100 V	99.99995 V	1 kHz	100.0000 V	$80 \cdot 10^{-6}$	0.0001 V	0.0222 V
100 V	100.00034 V	10 kHz	100.0003 V	$81 \cdot 10^{-6}$	0.0000 V	0.0222 V
100 V	100.00022 V	20 kHz	100.0002 V	$81 \cdot 10^{-6}$	0.0000 V	0.0222 V
100 V	99.99954 V	50 kHz	99.9995 V	$150 \cdot 10^{-6}$	0.0000 V	0.0372 V
100 V	99.99481 V	100 kHz	99.9948 V	$250 \cdot 10^{-6}$	0.0000 V	0.1222 V
1000 V	600.0008 V	40 Hz	600.001 V	$82 \cdot 10^{-6}$	0.000 V	0.281 V
1000 V	599.9993 V	100 Hz	599.999 V	$80 \cdot 10^{-6}$	0.000 V	0.261 V
1000 V	600.0046 V	1 kHz	600.005 V	$82 \cdot 10^{-6}$	0.000 V	0.261 V
1000 V	600.0095 V	10 kHz	600.010 V	$82 \cdot 10^{-6}$	0.000 V	0.381 V
1000 V	600.0117 V	20 kHz	600.012 V	$82 \cdot 10^{-6}$	0.000 V	0.381 V
1000 V	600.0282 V	50 kHz	600.028 V	$152 \cdot 10^{-6}$	0.000 V	0.741 V
1000 V	600.0835 V	100 kHz	600.084 V	$417 \cdot 10^{-6}$	0.000 V	1.821 V

Calibration Report

Certificate No.:

Page : 9 of 12

Calibration Results (Cont.)

Ohms Performance Tests

Test Conditions:

The 3458A was turned on at least 48-hours before calibration.

The 3458A's displayed internal temperature (TEMP?) was 38.0 °C

The 3458A was in the default/power-on state with the following exceptions:

Function: OHMF (For 4 wire ohms), OHM (For 2 wire ohms)

Ranges: Manual selection

OCOMP ON (Offset compensation, 10Ω to 10kΩ ranges)

NPLC 100 (Integration time 100 power line cycles)

The 3458A's front input terminals were used.

The Guard switch was in "Open" position.

The 3458A's "Guard" and "Lo" terminals were at Earth potential.

The 3458A was performed "AUTOCAL" prior to starting the measurements.

Ohms Zeroes Offset Test (MATH OFF, Shorting time > 5 min.)

UUC Range	Applied Value	UUC Reading	Uncertainty (±)	UUC Error	1Year Spec. (±)
10 Ω (2-W)	(Short) 0 Ω	0.00000 Ω	0.000020 Ω	0.00000 Ω	0.25007 Ω
10 Ω (4-W)	(Short) 0 Ω	0.00000 Ω	0.000020 Ω	0.00000 Ω	0.00007 Ω

4-Wire Ohms Gain Test

(Performed "MATH NULL" with the input terminals shorted, prior to starting calibration on each range)

UUC Range	Applied Value	UUC Reading	Uncertainty (±)	UUC Error	1Year Spec. (±)
10 Ω	1.0001337 Ω	1.00013 Ω	30 •10 ⁻⁶	0.00000 Ω	0.00007 Ω
10 Ω	9.999968 Ω	9.99997 Ω	26 •10 ⁻⁶	0.00000 Ω	0.00028 Ω
100 Ω	100.00015 Ω	100.00015 Ω	12 •10 ⁻⁶	0.00000 Ω	0.00231 Ω
1 kΩ	1.0000010 kΩ	1.0000010 kΩ	12 •10 ⁻⁶	0.0000000 kΩ	0.0000142 kΩ
10 kΩ	10.0001946 kΩ	10.000195 kΩ	8.1 •10 ⁻⁶	0.0000000 kΩ	0.000142 kΩ
100 kΩ	100.00072 kΩ	100.00072 kΩ	12 •10 ⁻⁶	0.00000 kΩ	0.00142 kΩ
1 MΩ	1.0000024 MΩ	1.0000024 MΩ	18 •10 ⁻⁶	0.0000000 MΩ	0.0000209 MΩ
10 MΩ	9.999629 MΩ	9.999629 MΩ	47 •10 ⁻⁶	0.0000000 MΩ	0.000703 MΩ
100 MΩ	99.903428 MΩ	99.90343 MΩ	80 •10 ⁻⁶	0.00000 MΩ	0.05125 MΩ

Calibration Report

Certificate No.:

Page : 10 of 12

Calibration Results (Cont.)

DC Current Performance Tests

Test Conditions:

The 3458A was turned on at least 48-hours before calibration.

The 3458A's displayed internal temperature (TEMP?) was 38.0 °C

The 3458A was in the default/power-on state with the following exceptions:

Function: DCI

Ranges: Manual selection

NPLC 100 (Integration time 100 power line cycles)

The 3458A's front input terminals were used.

The Guard switch was in "Open" position.

The 3458A's "Guard" and "Lo" terminals were at Earth potential.

The 3458A was performed "AUTOCAL" prior to starting the measurements.

DC Current Zeroes Offset Test (MATH OFF)

UUC Range	Applied Value	UUC Reading	Uncertainty (±)	UUC Error	1Year Spec. (±)
100 µA	(Open) 0 µA	0.00000 µA	0.00010 µA	0.00000 µA	0.00095 µA
1 mA	(Open) 0 mA	0.000000 mA	0.0010 µA	0.0000 µA	0.0065 µA
10 mA	(Open) 0 mA	0.000000 mA	0.010 µA	0.000 µA	0.065 µA
100 mA	(Open) 0 mA	0.00000 mA	0.10 µA	0.00 µA	0.65 µA
1 A	(Open) 0 A	0.000000 A	1.0 µA	0.0 µA	11.5 µA

DC Current Gain Test

(Performed "MATH NULL" with the input terminals opened, prior to starting calibration on each range)

UUC Range	Applied Value	UUC Reading	Uncertainty (±)	UUC Error	1Year Spec. (±)
100 µA	100.00043 µA	100.00043 µA	21 •10 ⁻⁶	0.00000 µA	0.00356 µA
100 µA	-100.00019 µA	-100.00019 µA	21 •10 ⁻⁶	0.00000 µA	0.00356 µA
1 mA	1.0000011 mA	1.0000011 mA	21 •10 ⁻⁶	0.0000000 mA	0.0000323 mA
1 mA	-0.9999994 mA	-0.9999994 mA	20 •10 ⁻⁶	0.0000000 mA	0.0000323 mA
10 mA	10.000015 mA	10.000015 mA	21 •10 ⁻⁶	0.000000 mA	0.000323 mA
10 mA	-10.000007 mA	-10.000007 mA	21 •10 ⁻⁶	0.000000 mA	0.000323 mA
100 mA	100.00120 mA	100.00120 mA	21 •10 ⁻⁶	0.00000 mA	0.00489 mA
100 mA	-100.00100 mA	-100.00100 mA	21 •10 ⁻⁶	0.00000 mA	0.00489 mA
1 A	1.0000532 A	1.0000532 A	36 •10 ⁻⁶	0.0000 mA	0.1349 mA
1 A	-1.0000445 A	-1.0000445 A	36 •10 ⁻⁶	0.0000 mA	0.1349 mA

Calibration Report

Certificate No.:

Page : 11 of 12

Calibration Results (Cont.)

AC Current Performance Test

Test Conditions:

The 3458A was turned on at least 48-hours before calibration.

The 3458A's displayed internal temperature (TEMP?) was 38.0 °C

The 3458A was in the default/power-on state with the following exceptions:

Function: ACI

Ranges: Manual selection

NPLC 100 (Integration time 100 power line cycles)

LFILTER ON (Low-pass filter ON)

The 3458A's front input terminals were used.

The Guard switch was in "Open" position.

The 3458A's "Guard" and "Lo" terminals were at Earth potential.

The 3458A was performed "AUTOCAL" prior to starting the measurements.

UUC Range	Applied Value	Frequency	UUC Reading	Uncertainty (\pm)	UUC Error	1Year Spec. (\pm)
100 μ A	100.0042 μ A	10 Hz	100.0042 μ A	$210 \cdot 10^{-6}$	0.0000 μ A	0.4305 μ A
100 μ A	100.0023 μ A	20 Hz	100.0023 μ A	$210 \cdot 10^{-6}$	0.0000 μ A	0.4305 μ A
100 μ A	100.0007 μ A	40 Hz	100.0007 μ A	$160 \cdot 10^{-6}$	0.0000 μ A	0.1805 μ A
100 μ A	99.9999 μ A	100 Hz	99.9999 μ A	$150 \cdot 10^{-6}$	0.0000 μ A	0.0905 μ A
100 μ A	99.9973 μ A	1 kHz	99.9973 μ A	$150 \cdot 10^{-6}$	0.0000 μ A	0.0905 μ A
100 μ A	100.0000 μ A	5 kHz	100.0000 μ A	$350 \cdot 10^{-6}$	0.0000 μ A	n/a
1 mA	1.000048 mA	10 Hz	1.000048 mA	$210 \cdot 10^{-6}$	0.000000 mA	0.004205 mA
1 mA	1.000027 mA	20 Hz	1.000027 mA	$210 \cdot 10^{-6}$	0.000000 mA	0.004205 mA
1 mA	1.000004 mA	40 Hz	1.000004 mA	$160 \cdot 10^{-6}$	0.000000 mA	0.001705 mA
1 mA	1.000003 mA	100 Hz	1.000003 mA	$160 \cdot 10^{-6}$	0.000000 mA	0.000805 mA
1 mA	0.999997 mA	1 kHz	0.999997 mA	$150 \cdot 10^{-6}$	0.000000 mA	0.000505 mA
1 mA	1.000000 mA	5 kHz	1.000000 mA	$250 \cdot 10^{-6}$	0.000000 mA	0.000505 mA
1 mA	1.000000 mA	10 kHz	1.000000 mA	$650 \cdot 10^{-6}$	0.000000 mA	0.000805 mA
10 mA	10.00036 mA	10 Hz	10.00036 mA	$160 \cdot 10^{-6}$	0.00000 mA	0.04205 mA
10 mA	10.00010 mA	20 Hz	10.00010 mA	$160 \cdot 10^{-6}$	0.00000 mA	0.04205 mA
10 mA	10.00001 mA	40 Hz	10.00001 mA	$160 \cdot 10^{-6}$	0.00000 mA	0.01705 mA
10 mA	10.00002 mA	100 Hz	10.00002 mA	$160 \cdot 10^{-6}$	0.00000 mA	0.00805 mA
10 mA	9.99994 mA	1 kHz	9.99994 mA	$150 \cdot 10^{-6}$	0.00000 mA	0.00505 mA
10 mA	10.00000 mA	5 kHz	10.00000 mA	$250 \cdot 10^{-6}$	0.00000 mA	0.00505 mA
10 mA	10.00000 mA	10 kHz	10.00000 mA	$640 \cdot 10^{-6}$	0.00000 mA	0.00805 mA

Calibration Report

Certificate No.:

Page : 12 of 12

Calibration Results (Cont.)

AC Current Performance Test

Cont.

UUC Range	Applied Value	Frequency	UUC Reading	Uncertainty (\pm)	UUC Error	1Year Spec. (\pm)
100 mA	100.0042 mA	10 Hz	100.0042 mA	$160 \cdot 10^{-6}$	0.0000 mA	0.4205 mA
100 mA	100.0024 mA	20 Hz	100.0024 mA	$160 \cdot 10^{-6}$	0.0000 mA	0.4205 mA
100 mA	100.0018 mA	40 Hz	100.0018 mA	$160 \cdot 10^{-6}$	0.0000 mA	0.1705 mA
100 mA	100.0015 mA	100 Hz	100.0015 mA	$160 \cdot 10^{-6}$	0.0000 mA	0.0805 mA
100 mA	100.0009 mA	1 kHz	100.0009 mA	$160 \cdot 10^{-6}$	0.0000 mA	0.0505 mA
100 mA	100.0000 mA	5 kHz	100.0000 mA	$250 \cdot 10^{-6}$	0.0000 mA	0.0505 mA
100 mA	100.0000 mA	10 kHz	100.0000 mA	$640 \cdot 10^{-6}$	0.0000 mA	0.0805 mA
1 A	1.000089 A	10 Hz	1.000089 A	$160 \cdot 10^{-6}$	0.000000 A	0.004205 A
1 A	1.000072 A	20 Hz	1.000072 A	$160 \cdot 10^{-6}$	0.000000 A	0.004205 A
1 A	1.000070 A	40 Hz	1.000070 A	$160 \cdot 10^{-6}$	0.000000 A	0.001805 A
1 A	1.000070 A	100 Hz	1.000070 A	$160 \cdot 10^{-6}$	0.000000 A	0.001005 A
1 A	1.000064 A	1 kHz	1.000064 A	$160 \cdot 10^{-6}$	0.000000 A	0.001205 A
1 A	1.000000 A	5 kHz	1.000000 A	$190 \cdot 10^{-6}$	0.000000 A	0.001205 A
1 A	1.000000 A	10 kHz	1.000000 A	$190 \cdot 10^{-6}$	0.000000 A	0.003205 A

Frequency Performance Test

Test Conditions:

The 3458A was turned on at least 48-hours before calibration.

The 3458A's displayed internal temperature (TEMP?) was 38.0 °C

The 3458A was in the default/power-on state with the following exceptions:

Function: FREQ

Fsource ACDCV

Level 0, DC

LFILTER ON @ 1Hz, 10MHz OFF (Low-pass filter)

The 3458A's front input terminals were used.

The Guard switch was in "To Lo" position.

The 3458A was performed "AUTOCAL" prior to starting the measurements.

Test Voltage	Applied Value	UUC Reading	Uncertainty (\pm)	UUC Error	1Year Spec. (\pm)
1 Vp-p	1 Hz	1.000000 Hz	$10 \cdot 10^{-6}$	0.000000 Hz	0.000500 Hz
1 Vp-p	10 MHz	10.000000 MHz	$6.0 \cdot 10^{-6}$	0.000000 MHz	0.00100 MHz

-000-